

Title	Excitation functions for the production of ^{81}Rb - ^{81}Kr via the $^{79}\text{Br}(\alpha, 2n)^{81}\text{Rb}$ and the $^{81}\text{Br}(^3\text{He}, 3n)^{81}\text{Rb}$ reactions
Sub Title	
Author	本間, 義夫(Honma, Yoshio) 倉田, 恭子(Kurata, Kyoko)
Publisher	共立薬科大学
Publication year	1978
Jtitle	共立薬科大学研究年報 (The annual report of the Kyoritsu College of Pharmacy). No.23 (1978.) ,p.124- 124
JaLC DOI	
Abstract	
Notes	抄録
Genre	Technical Report
URL	https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000023-0124

慶應義塾大学学術情報リポジトリ(KOARA)に掲載されているコンテンツの著作権は、それぞれの著作者、学会または出版社/発行者に帰属し、その権利は著作権法によって保護されています。引用にあたっては、著作権法を遵守してご利用ください。

The copyrights of content available on the Keio Associated Repository of Academic resources (KOARA) belong to the respective authors, academic societies, or publishers/issuers, and these rights are protected by the Japanese Copyright Act. When quoting the content, please follow the Japanese copyright act.

**Excitation Functions for the Production of ^{81}Rb - ^{81m}Kr via the ^{79}Br
(α , 2n) ^{81}Rb and the ^{81}Br (^3He , 3n) ^{81}Rb Reactions ***

YOSHIO HOMMA and KYOKO KURATA

本間義夫, 倉田恭子

Krypton-81m is said to be an ideal radionuclide for scanning in a large variety of studies of nuclear medicine: To date, most of the ^{81}Rb - ^{81m}Kr used in nuclear medical applications has been produced by the following nuclear reactions: ^{85}Rb (p, 5n) ^{81}Sr - ^{81}Rb and ^{79}Br (α , 2n) ^{81}Rb . The limitation of the former nuclear reaction is that the required particle energies (68 MeV) are not attainable with the compact medical cyclotron.

Earlier data concerning the latter reaction have been restricted to the measurement of ^{81}Rb - ^{81m}Kr and the by-product nuclides such as ^{82m}Rb , ^{83}Rb and ^{84}Rb . In order to determine the optimum irradiation conditions to maximize the yield of ^{81m}Kr and minimize the yields of other radionuclides of krypton, excitation functions and thick target yield curves have been measured by means of the stacked foil technique for α and ^3He reactions producing ^{81}Rb - ^{81m}Kr , ^{82m}Rb , ^{83}Rb - ^{83m}Kr , ^{84}Rb , ^{77}Br and ^{79}Kr , both from potassium bromide.

The ^{79}Br (α , 2n) ^{81}Rb and ^{81}Br (^3He , 3n) ^{81}Rb reactions have cross section peaks of 300 mb and 320 mb at 28 MeV and 29 MeV, respectively.

The ^{79}Br (α , 2n) ^{81}Rb - ^{81m}Kr reaction will provide satisfactory yield 2 mCi/ μAh for most nuclear medical applications. However, the ^{83m}Kr contamination limits the products usefulness in high-resolution nuclear medical applications and the need for 40 MeV particles restricts the production to higher energy cyclotrons.

On the other hand, the ^{81}Br (^3He , 3n) ^{81}Rb - ^{81m}Kr reaction is particularly advantageous since the ^{83m}Kr contamination can be reduced to an almost negligible. Under our experimental conditions (30 MeV ^3He , 1 μA for an hour) about 2 mCi ^{81m}Kr can be produced. This yield is comparable to that for the α reaction. Moreover, the compact cyclotron cannot accelerate α particles to 35 MeV to 40 MeV necessary to produce sufficient ^{81}Rb - ^{81m}Kr by the ^{79}Br (α , 2n) ^{81}Rb reaction, whereas the ^{81}Br (^3He , 3n) ^{81}Rb reaction needs 30 MeV ^3He particles, which is available from some compact cyclotrons. This shown that the ^3He bombardment is especially useful if one is limited to use of ^{81m}Kr produced solely by a compact cyclotron, although it has not been reported in the past.

本報告は Int. J. appl. Radiat. Isotopes. に発表。(in press).