

Title	Determination of osalimid in plasma by high-performance liquid chromatography
Sub Title	
Author	定永, 恒雄(Sadanaga, Tsuneo) 正田, 興造( Hikida, Kozo) 為藤, 勝信( Tameto, Katsunobu) 杉本, 淳( Sugimoto, Kiyoshi) 中西, 美智夫( Nakanishi, Michio) 松島, 美一( Matsushima, Yoshikazu) 大倉, 洋甫( Okura, Yosuke)
Publisher	共立薬科大学
Publication year	1982
Jtitle	共立薬科大学研究年報 (The annual report of the Kyoritsu College of Pharmacy). No.27 (1982. ) ,p.77- 77
JaLC DOI	
Abstract	
Notes	抄録
Genre	Technical Report
URL	<a href="https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000027-0077">https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000027-0077</a>

慶應義塾大学学術情報リポジトリ(KOARA)に掲載されているコンテンツの著作権は、それぞれの著作者、学会または出版社/発行者に帰属し、その権利は著作権法によって保護されています。引用にあたっては、著作権法を遵守してご利用ください。

The copyrights of content available on the Keio Associated Repository of Academic resources (KOARA) belong to the respective authors, academic societies, or publishers/issuers, and these rights are protected by the Japanese Copyright Act. When quoting the content, please follow the Japanese copyright act.

## Determination of Osalmid in Plasma by High- Performance Liquid Chromatography\*

TSUNEO SADANAGA\*\*, KOZO HIKIDA\*\*, KATSUNOBU TAMETO\*\*, KIYOSHI  
SUGIMOTO\*\*, MICHIO NAKANISHI\*\*, YOSHIKAZU MATSUSHIMA,  
and YOSUKE OHKURA\*\*\*

定永恒雄,\*\* 疋田興造,\*\* 為藤勝信,\*\* 杉本 淳,\*\*  
中西美智夫,\*\* 松島美一, 大倉洋甫\*\*\*

Osalmid, 4'-hydroxysalicylanilide, is currently used as a choleretic drug. A rapid and sensitive method of assay for osalmid was required in the course of investigations on its absorption, distribution, metabolism, and excretion. High-performance liquid chromatography (HPLC) was found to meet the requirement.

Plasma mixed with phosphate buffer was extracted by isopropyl ether containing 5% isoamyl alcohol. The organic layer was separated and shaken with aqueous NaOH. The aqueous layer was separated, acidified with aqueous HCl, and extracted with 10% isoamyl alcohol-n-butyl acetate containing 2,7-dihydroxynaphthalene as a reference standard. After the complete removal of the aqueous layer, a portion of the organic layer was injected into the chromatographic system.

A Hitachi liquid chromatograph system equipped with UV monitor operated at 300 nm and a LiChrosorb Si-100 column were used. The mobile phase was dichloromethane-n-hexane-methanol-5% acetic acid (66:30:4:0.35, v/v). The ratio of the peak height of osalmid to that of the reference standard was plotted against the amount of osalmid added to the drug-free plasma. The calibration curve thus obtained was linear and passed through the origin.

The method permits the accurate determination of osalmid in plasma at concentrations as low as 9 ng/ml. The precision and recoveries were quite satisfactory. By the present method, plasma concentrations of osalmid at various times after oral administration to dogs were investigated.

\* 本報告は *J. Chromatogr.*, 223, 243~246 (1981) に発表.

\*\* 吉富製薬.

\*\*\* 九州大学薬学部.