Title	Effect of air luminescence counts on determination of ²²² Rn by liquid scintillation counting
Sub Title	
Author	本間, 義夫(Honma, Yoshio) 村瀬, 裕子(Murase, Yuko) 滝上, 誠(Takiue, Makoto)
Publisher	共立薬科大学
Publication year	1987
Jtitle	共立薬科大学研究年報 (The annual report of the Kyoritsu College of Pharmacy). No.32 (1987.) ,p.84- 84
JaLC DOI	
Abstract	
Notes	抄録
Genre	Technical Report
URL	https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000032- 0084

慶應義塾大学学術情報リポジトリ(KOARA)に掲載されているコンテンツの著作権は、それぞれの著作者、学会または出版社/発行者に帰属し、その権利は著作権法によって 保護されています。引用にあたっては、著作権法を遵守してご利用ください。

The copyrights of content available on the KeiO Associated Repository of Academic resources (KOARA) belong to the respective authors, academic societies, or publishers/issuers, and these rights are protected by the Japanese Copyright Act. When quoting the content, please follow the Japanese copyright act.

Effect of Air Luminescence Counts on Determination of ²²²Rn by Liquid Scintillation Counting*

Yoshio Homma, Yuko Murase and Makoto Takiue**

本間義夫, 村瀬裕子, 滝上 誠**

The advantages of liquid scintillation counting for the determination of disintegration rates of ²²²Rn and the daughters have been demonstrated by a number of investigators. The use of integral counting technique and the fact that ²²²Rn is appreciably soluble in liquid scintillator make the method particularly useful. However, in the course of a spectral study concerning the relative pulse-height distributions of ²²²Rn and the daughters in liquid scintillator, we found that, when the counting vial is not full of liquid scintillator, the ²²²Rn in the space above the liquid scintillator causes air luminescence counts, which has been hitherto neglected by the conventional liquid scintillation counting of ²²²Rn.

In order to obtain the accurate disintegration rates of ²²²Rn and the daughters, it is necessary to determine the disintegration rates that are only due to the results of excitation from ²²²Rn and the daughters dissolved in the liquid scintillator. Since the air luminescence spectra are superimposed upon the beta continuum of ²¹⁴Pb and ²¹⁴Bi, the disintegration rates of ²²²Rn and the daughters dissolved in the scintillator solution were obtained by subtracting the air luminescence spectrum from the corresponding total pulse-height spectrum.

The spectral analysis of the ²²²Rn sample indicate that ²²²Rn and the daughters in the gaseous space do produce the air luminescence counts and lead to erroneously high ²²²Rn activities, unless those air luminescence counts are subtracted from the observed counts. The estimated errors due to the air luminescence are : 12.9% for 8 ml, 6.4% for 12 ml, 2.8% for 16 ml, and 0.3% for 20 ml.

The authors have reached opinion that the most practical method for minimizing the error caused by the air luminescence is to measure 222 Rn in the scintillator volume of not less than 20 ml. However, it should be noted that there is a appreciable fall in counting efficiency between 15 to 20 ml of liquid scintillator volume. A detailed investigation concerning the relationship between the volume of scintillator solution and the counting efficiency, which is underway and will be published separately, showed that this fall in counting efficiency could be corrected using integral counting techniques with precision better than 1.2%.

^{*} 本報告は Appl. Radiat. Isot. に発表. (in Press).

^{**} 慈恵会医科大学