Title	Determination of air luminescence spectra for alpha-emitters with liquid scintillation spectrometers
Sub Title	
Author	村瀬, 裕子(Murase, Yuko) 本間, 義夫(Honma, Yoshio) 滝上, 誠(Takiue, Makoto) 油井, 多丸(Aburai, Tamaru)
Publisher	共立薬科大学
Publication year	1988
Jtitle	共立薬科大学研究年報 (The annual report of the Kyoritsu College of Pharmacy). No.33 (1988.),p.173- 173
JaLC DOI	
Abstract	
Notes	抄録
Genre	Technical Report
URL	https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000033-0173

慶應義塾大学学術情報リポジトリ(KOARA)に掲載されているコンテンツの著作権は、それぞれの著作者、学会または出版社/発行者に帰属し、その権利は著作権法によって保護されています。引用にあたっては、著作権法を遵守してご利用ください。

The copyrights of content available on the Keio Associated Repository of Academic resources (KOARA) belong to the respective authors, academic societies, or publishers/issuers, and these rights are protected by the Japanese Copyright Act. When quoting the content, please follow the Japanese copyright act.

Determination of Air Luminescence Spectra for Alpha-Emitters with Liquid Scintillation Spectrometers

Yuko MURASE, Yoshio HOMMA, Makoto TAKIUE**, and Tamaru ABURAI***

村瀬裕子, 本間義夫, 滝上 誠**, 油井多丸***

Air luminescence spectra have been measured for ^{210}Po , ^{238}U and ^{241}Am with two commercially available liquid scintillation spectrometers. Since the measured end-point of the air luminescence spectra is approx. 18 keV, they can be determined by liquid scintillation counting in the ^3H detection region. The average counting efficiencies measured with both counting systems were found to be 33.3% and 33.7%, respectively. The air luminescence effect on liquid scintillation counting of gaseous α -emitters is mentioned.

* 本報告は *Appl. Radiat. Isot.* (in press)

** 慈恵会医科大学

*** 日本原子力研究所ラジオアイソトープ・原子炉研修所.