

Title	High-performance liquid chromatography of antipsychotic drugs with a porous glass ODS
Sub Title	
Author	藤江, 忠雄(Fujie, Tadao) 金子, 明子(Kaneko, Akiko) 金子, 和子(Kaneko, Kazuko) 杉本, 圭子(Sugimoto, Keiko) 高井, 信治(Takai, Nobuharu)
Publisher	共立薬科大学
Publication year	1988
Jtitle	共立薬科大学研究年報 (The annual report of the Kyoritsu College of Pharmacy). No.33 (1988.) ,p.179- 179
JaLC DOI	
Abstract	
Notes	抄録
Genre	Technical Report
URL	https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000033-0179

慶應義塾大学学術情報リポジトリ(KOARA)に掲載されているコンテンツの著作権は、それぞれの著作者、学会または出版社/発行者に帰属し、その権利は著作権法によって保護されています。引用にあたっては、著作権法を遵守してご利用ください。

The copyrights of content available on the KeiO Associated Repository of Academic resources (KOARA) belong to the respective authors, academic societies, or publishers/issuers, and these rights are protected by the Japanese Copyright Act. When quoting the content, please follow the Japanese copyright act.

High-Performance Liquid Chromatography of Antipsychotic Drugs with a Porous Glass ODS

Tadao FUJIE, Akiko KANEKO, Kazuko KANEKO,
Keiko SUGIMOTO, and Nobuharu TAKAI*

藤江忠雄, 金子明子, 金子和子, 杉本圭子, 高井信治*

Liquid chromatography has been rapidly developed recently, especially in capability of good separation in shorter time. Since this instrument is still expensive even today, it is recommended to obtain more information out of a single measurement, that is to say, if more rapid analysis for separation could be achieved, more information would be obtained in a restricted time. From this point of view, the following studies were carried out.

One of the approaches to meet the above requirement is to use the larger column for the increase of flow rate. However, it is not considered to be a economical method, since the large amount of eluent is required. In this studies, therefore, attention was made to improve the packing materials while maintaining the flow rate of eluent kept the same as normal flow rate. We have already tried to regulate the amount of chemical modification on the surface of the packing materials, to vary the carbon chain length, and to change the pore structure of the packing materials. All of the above approaches were found effective for more rapid analysis.

In the present studies, we report that, the good chromatograms were obtained to meet with the above requirements, when porous glass is chemically modified.

* 本報告は *J. Life Support Techn. Soc.*, 2 (1), 134-136 (1987) に発表.

* 東京大学生産技術研究所