

Title	Application of glass ODS for the HPLC analysis of estrone 17- $\beta$ -estradiol and estriol
Sub Title	
Author	小股, 泰子(Komata, Yasuko) 金子, 明子(Kaneko, Akiko) 藤江, 忠雄(Fujie, Tadao) 高井, 信治(Takai, Nobuharu)
Publisher	共立薬科大学
Publication year	1988
Jtitle	共立薬科大学研究年報 (The annual report of the Kyoritsu College of Pharmacy). No.33 (1988. ) ,p.180- 180
JaLC DOI	
Abstract	
Notes	抄録
Genre	Technical Report
URL	<a href="https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000033-0180">https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000033-0180</a>

慶應義塾大学学術情報リポジトリ(KOARA)に掲載されているコンテンツの著作権は、それぞれの著作者、学会または出版社/発行者に帰属し、その権利は著作権法によって保護されています。引用にあたっては、著作権法を遵守してご利用ください。

The copyrights of content available on the KeiO Associated Repository of Academic resources (KOARA) belong to the respective authors, academic societies, or publishers/issuers, and these rights are protected by the Japanese Copyright Act. When quoting the content, please follow the Japanese copyright act.

## Application of Glass ODS for the HPLC Analysis of Estrone 17- $\beta$ -Estradiol and Estriol

Yasuko KOMATA, Akiko KANEKO, Tadao FUJIE, and Nobuharu TAKAI\*

小股泰子, 金子明子, 藤江忠雄, 高井信治\*

Reversed phase high-performance liquid chromatography (HPLC) with an ODS porous glass column (HITACHI GEL #3161) was used to assay estrone, 17- $\beta$ -estradiol and estriol. Using acetonitrile-water (30/70) as a mobile phase at a flow rate of 1 ml/min, the three steroids were analysed in 5 min. The results demonstrate that the present packing materials was much advantageous to analyse these three steroids.

---

本報告は *J. Life Support Techn. Soc.*, 2 (1), 137-139 (1987) に発表.

\* 東京大学生産技術研究所