

Title	An improved extractor for dual-flow extraction
Sub Title	
Author	西沢, 秀幸(Nishizawa, Hideyuki) 渡辺, 右佳子(Watanabe, Yukako) 沖村, 早希子(Okimura, Sakiko) 阿部, 芳廣(Abe, Yoshihiro)
Publisher	共立薬科大学
Publication year	1989
Jtitle	共立薬科大学研究年報 (The annual report of the Kyoritsu College of Pharmacy). No.34 (1989.) ,p.66- 66
JaLC DOI	
Abstract	
Notes	抄録
Genre	Technical Report
URL	https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000034-0066

慶應義塾大学学術情報リポジトリ(KOARA)に掲載されているコンテンツの著作権は、それぞれの著作者、学会または出版社/発行者に帰属し、その権利は著作権法によって保護されています。引用にあたっては、著作権法を遵守してご利用ください。

The copyrights of content available on the KeiO Associated Repository of Academic resources (KOARA) belong to the respective authors, academic societies, or publishers/issuers, and these rights are protected by the Japanese Copyright Act. When quoting the content, please follow the Japanese copyright act.

An Improved Extractor for Dual-Flow Extraction*

Hideyuki NISHIZAWA, Yukako WATANABE,
Sakiko OKIMURA and Yoshihiro ABE

西沢秀幸, 渡辺右佳子, 沖村早希子, 阿部芳廣

An improved apparatus for dual-flow countercurrent extraction was constructed. Further, a simple method to evaluate the extractor was developed, by which the effective plate number could be calculated by a couple of measurements of fractional recoveries. To find the optimum operation condition, an apparatus with a shorter column (10 cm in length) was prepared, the efficiency of which was evaluated as being about one plate per cm of the column length by the method. It was also shown that the distribution constant of a compound could be calculated from the fractional recoveries measured by dual-flow countercurrent extraction. The distribution constants, thus obtained, agreed with values determined independently by liquid-liquid chromatography, which settled theoretical correlations between chromatography and extraction.

* 本報告は *Anal. Sci.*, 5 (3), 339-343 (1989) に発表.