Title	Reliability of the activity determined by the quenching correction method for two groups of emulsion scintillators
Sub Title	
Author	滝上, 誠(Takiue, Makoto) 藤井, 張生(Fujii, Haruo) 本間, 義夫(Honma, Yoshio)
Publisher	共立薬科大学
Publication year	1989
Jtitle	共立薬科大学研究年報 (The annual report of the Kyoritsu College of Pharmacy). No.34 (1989.) ,p.95- 95
JaLC DOI	
Abstract	
Notes	抄録
Genre	Technical Report
URL	https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000034-0095

慶應義塾大学学術情報リポジトリ(KOARA)に掲載されているコンテンツの著作権は、それぞれの著作者、学会または出版社/発行者に帰属し、その権利は著作権法によって保護されています。引用にあたっては、著作権法を遵守してご利用ください。

The copyrights of content available on the Keio Associated Repository of Academic resources (KOARA) belong to the respective authors, academic societies, or publishers/issuers, and these rights are protected by the Japanese Copyright Act. When quoting the content, please follow the Japanese copyright act.

Reliability of the Activity Determined by the Quenching Correction Method for Two Groups of Emulsion Scintillators*

Makoto TAKIUE**, Haruo FUJII*** and Yoshio HOMMA

滝上 誠**, 藤井張生***, 本間義夫

The reliability of the activity, determined using conventional quenching correction methods, has been studied for two groups of emulsion scintillators: (1) scintillators with three distinct water dilution criteria; and (2) scintillators forming a miscible solution with up to 50% water content. Whilst for the first group of scintillators the error of the activity measured becomes frequently larger with increased water content, the measurements with the second group of scintillators fit the quenching correction curve.

* 本報告は *Appl. Radiat. Isot.* に発表. (in press)

** 慈恵会医科大学

*** 東京医科歯科大学