Title	Determination of hydroperoxides with the use of immobilized glutathione peroxidase in a flow system
Sub Title	
Author	森, 久和(Mori, Hisakazu) 小暮, 真美絵(Kogure, Mamie) 隈部(山品), 恭子(Kumabe(Yamashina), Kyoko)
Publisher	共立薬科大学
Publication year	1992
Jtitle	共立薬科大学研究年報 (The annual report of the Kyoritsu College of Pharmacy). No.37 (1992.),p.62- 62
JaLC DOI	
Abstract	
Notes	抄録
Genre	Technical Report
URL	https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000037-0062

慶應義塾大学学術情報リポジトリ(KOARA)に掲載されているコンテンツの著作権は、それぞれの著作者、学会または出版社/発行者に帰属し、その権利は著作権法によって保護されています。引用にあたっては、著作権法を遵守してご利用ください。

The copyrights of content available on the KeiO Associated Repository of Academic resources (KOARA) belong to the respective authors, academic societies, or publishers/issuers, and these rights are protected by the Japanese Copyright Act. When quoting the content, please follow the Japanese copyright act.

Determination of Hydroperoxides with the Use of Immobilized Glutathione Peroxidase in a Flow System*

Hisakazu MORI, Mamie KOGURE and Kyoko KUMABE (nee YAMASHINA)

森 久和, 小暮真美絵, 隅部(旧姓山品) 恭子

Various hydroperoxides were determined in a flow system with the use of glutathione peroxidase immobilized on N-(2-aminoethyl)-3-aminopropyl glass. In this system, the glutathione diminished by the reaction with hydroperoxide was monitored through its reaction with 5,5'-dithio-bis-(2-nitrobenzoic acid). The optimum pH for this immobilized enzyme was found to be 7.8. The rates of three carrier solutions were examined to obtain a favorable peak. The calibration curves for hydrogen peroxide, tert-butyl hydroperoxide, cumene hydroperoxide, linoleate hydroperoxide and linolenate hydroperoxide showed high linearity over the sample concentration range from not more than 10 to 100 μM . The detection limits for the above hydroperoxides were in the range from 0.2 to 0.5 nmol. The determination of hydroperoxides could be performed in 4 min.

* 本報告は *Anal. Lett.*, **25** (9), 1643-1656 (1992) に発表.