Title	Determination of inorganic phosphate by the use of immobilized enzymes in a FIA system
Sub Title	
Author	森, 久和(Mori, Hisakazu) 小暮, 真美絵(Kogure, Mamie) 川俣, 寿美代(Kawamata, Sumiyo) 永本, 昭子(Nagamoto, Akiko) 山本, 晴彦(Yamamoto, Haruhiko)
Publisher	共立薬科大学
Publication year	1993
Jtitle	共立薬科大学研究年報 (The annual report of the Kyoritsu College of Pharmacy). No.38 (1993.),p.70- 70
JaLC DOI	
Abstract	
Notes	抄録
Genre	Technical Report
URL	https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000038-0070

慶應義塾大学学術情報リポジトリ(KOARA)に掲載されているコンテンツの著作権は、それぞれの著作者、学会または出版社/発行者に帰属し、その権利は著作権法によって保護されています。引用にあたっては、著作権法を遵守してご利用ください。

The copyrights of content available on the Keio Associated Repository of Academic resources (KOARA) belong to the respective authors, academic societies, or publishers/issuers, and these rights are protected by the Japanese Copyright Act. When quoting the content, please follow the Japanese copyright act.

Determination of Inorganic Phosphate by the Use of Immobilized Enzymes in a FIA System*

Hisakazu MORI, Mamie KOGURE, Sumiyo KAWAMATA, Akiko NAGAMOTO
and Haruhiko YAMAMOTO**

森 久和, 小暮真美絵, 川俣寿美代, 永本昭子, 山本晴彦**

Inorganic phosphate was fluorometrically determined in the FIA system with a column containing co-immobilized enzymes for successive reactions to generate NADPH from inorganic phosphate. The enzymes used were sucrose phosphorylase, phosphoglucosaminidase and glucose-6-phosphate dehydrogenase, and they were co-immobilized on amino-propyl glass. Assay conditions such as flow rate of a carrier solution and column temperature were investigated for optimization. Several buffers of different pHs were examined as a carrier solution to find Tris buffer (pH 8.0) most favorable. Under the conditions, the calibration curve for the determination of inorganic phosphate (50 μ l injection) showed a linear relationship in the concentration range of 0.2–100 μ M. The detection limit was 0.1 μ M, that is, 5 pmol. The relative standard deviation of the peak area for 10 μ M sample was 4.1% ($n = 5$). The analysis of a sample completed in 10 min. This method was applied to the determination of phosphorus content in the calf thymus DNA. The content obtained was coincident with that obtained by the molybdenum blue method.

* 本報告は *Anal. Lett.* に発表。(in press).

** 神奈川大学理学部