Title	Development of an optical-fibre sensor using a functional membrane
Sub Title	
Author	高井, 信治(Takai, Nobuharu) 平井, 利志(Hirai, Toshiyuki) 佐久間, 一郎(Sakuma, Ichiro) 福井, 康裕(Fukui, Yasuhiro) 金子, 明子(Kaneko, Akiko) 藤江, 忠雄(Fujie, Tadao)
Publisher	共立薬科大学
Publication year	1993
Jtitle	共立薬科大学研究年報 (The annual report of the Kyoritsu College of Pharmacy). No.38 (1993.),p.87- 87
JaLC DOI	
Abstract	
Notes	抄録
Genre	Technical Report
URL	https://koara.lib.keio.ac.jp/xoonips/modules/xoonips/detail.php?koara_id=AN00062898-00000038-0087

慶應義塾大学学術情報リポジトリ(KOARA)に掲載されているコンテンツの著作権は、それぞれの著作者、学会または出版社/発行者に帰属し、その権利は著作権法によって保護されています。引用にあたっては、著作権法を遵守してご利用ください。

The copyrights of content available on the Keio Associated Repository of Academic resources (KOARA) belong to the respective authors, academic societies, or publishers/issuers, and these rights are protected by the Japanese Copyright Act. When quoting the content, please follow the Japanese copyright act.

Development of an Optical-fibre Sensor Using a Functional Membrane*

Nobuharu TAKAI**, Toshiyuki HIRAI***, Ichiro SAKUMA***, Yasuhiro FUKUI***,
Akiko KANEKO and Tadao FUJIE

高井信治**, 平井利志**, 佐久間一郎***, 福井康裕***, 金子明子, 藤江忠雄

A new functional membrane which changes its reflectance spectra according to the pH of the test solution is described. A chloromethylstyrene (CMS)/methylmethacrylate (MMA) copolymer was chosen as the base for the immobilized reagent. A basic ion-exchanger was then added the CMS/MMA copolymer. Bromothymol Blue (BTB) was then immobilized on the copolymer. A functional membrane could then be fabricated in various shapes since chemical modification of CMS/MMA copolymer could be performed before casting. The fabricated pH sensitive functional membrane showed a maximum reflectance change at the wavelength of 540 nm for pH changes from 4.0 to 9.0. Dissolution of BTB from the CMS/MMA copolymer was minimal so that good reproducibility of the reflectance change was obtained.

* 本報告は *Sensors and Actuators B*. 13—14, 427—428 (1993) に発表。

** 東大・生産研

*** 東京電機大